



# ZPA NDIR/O<sub>2</sub> Multichannel Analyzer



The ZPA analyzer uses NDIR technology to measure CO, CO<sub>2</sub>, CH<sub>4</sub>, SO<sub>2</sub>, NO and paramagnetic or fuel cell for oxygen

## Features

- Multi-Component analyzer—Up to Four NDIR channels Plus Oxygen
- Measures NDIR gases from low ppm Up to 100% full scale, oxygen from 1 to 100%
- Virtually unaffected by moisture by moisture interference
- Measures oxygen via either paramagnetic or fuel cell
- Outputs: Voltage, Current, RS-485
- CE Mark
- Compact size

## Applications

- Continuous emission monitoring (CEMS)
- Process gas analysis
- Greenhouse gases
- Stack testing
- Gas purity
- Research

## Options

- Oxygen only version
- Fault Alarm
- External NOx converter
- High/low alarm
- Automatic Calibration
- Pump Pak II sample pump

## California Analytical Instruments

1312 West Grove Avenue, Orange, CA 92865 • Phone: 714-974-5560 • Fax: 714-921-2531  
www.gasanalyzers.com



# ZPA NDIR/O<sub>2</sub> Multichannel Analyzer

## Method of Operation

The ZPA NDIR technology is based on the infrared absorption characteristics of gases. Using a single infrared beam to measure gas concentrations, this analyzer produces highly stable and reliable results. A single infrared light beam is modulated by a chopper system and passed through a sample cell of predetermined length containing the gas sample to be analyzed. As the beam passes through the cell, the sample gas absorbs some of its energy. The attenuated beam (transmittance) emerges from the cell and is introduced to the front chamber of a two-chamber infrared microflow detector.

The detector is filled with the gas component of interest and consequently the beam experiences further energy absorption. This absorption process increases the pressure in both of the chambers. The differential pressure between the front and rear chambers of the detector causes a slight gas flow between the two chambers. This flow is detected by a mass-flow sensor and is converted into an output signal.

The oxygen channel of the ZPA utilizes either the paramagnetic or fuel cell method to determine the percent level of oxygen contained in the sample gas.

## Specifications

**IR Analysis Method:** Non-Dispersive Infrared (NDIR)  
**NDIR Components:** CO / CO<sub>2</sub> / CH<sub>4</sub> / SO<sub>2</sub>/NO  
**Detector Type:** Microflow  
**NDIR Ranges:**

Gas	Minimum	Maximum
CO	0-200 ppm	0-100%
CO <sub>2</sub>	0-200 ppm	0-100%
CH <sub>4</sub>	0-500 ppm	0-100%
SO <sub>2</sub>	0-200 ppm	0-10%
NO	0-200 ppm	0-5,000 ppm

**Range Ratio:** 10:1

**Oxygen Analysis Method:** Paramagnetic or Fuel Cell

**O<sub>2</sub> Ranges:** 0-5% or 0-25% Full Scale

**Response Time (IR):** 90% of Full Scale within 30 seconds

\*\*Depending on Cell Length, Flow Rate, and Time Constant

**IR Sample Cell:** Stainless Steel w/ Replaceable gold cell liner

**Resolution:** Typically 0.1% of Full Scale

**Repeatability:** Better than 0.5% of Full Scale

**Linearity:** Better than 1.0% of Full Scale of Factory

**Noise:** Less than 1% of Full scale Range

**Zero & Span Drift:** Less than 2% of Full Scale per week

**Zero & Span Adjustment:** Via front panel

**Sample Flow Rate:** 0.5 LPM  
**Purge Gas Flow Rate:** 1 LPM  
**Outputs available:** USB, RS485 (MODBUS protocol), 0-1V / 4-20mA

**Digital Inputs:** 9 maximum optically isolated signals for range switching, begin auto calibration, output signal hold, reset average value

**Digital Outputs:** 15 relays maximum, each is 1 form C for range ID, instrument failure, cal failure, cal in progress, high/low alarm limits, solenoid valve activation, external pump on/off

**Display:** Back lit LCD

**Sample Temperature:** Up to 50 C, Non-condensing

**Ambient Temperature:** -5 to 50 C

**Ambient Humidity:** Less than 90% RH (Non-condensing)

**Fittings:** ¼ inch NPT

**Power Requirements:** 100 to 240 VAC, 50/60 Hz, 100 VA

**Dimensions:** 5¼"Hx19"Wx15"D

**Weight:** Approximately 18 lbs. (Depending on configuration)

Specifications subject to change without notice.

## California Analytical Instruments

1312 West Grove Avenue, Orange, CA 92865 • Phone: 714-974-5560 • Fax: 714-921-2531  
www.gasanalyzers.com